Kundeninformation zur Verlängerung der MDR-Zulassung 28.11.2023

Zertifikatsnummer: G1 036336 0054 Rev. 03

Seit dem 20.03.2023 gilt eine Verlängerung der bisherigen Übergangsfrist zur Umsetzung der MDR. Diese Verlängerung wurde von der EU beschlossen, da die Umstellung weder von den Herstellern noch von den Benannten Stellen im vorgesehenen Zeitraum zu leisten war.

Sind nun Zertifikate von Herstellern, die noch nicht MDR-zertifiziert sind, abgelaufen, ist eine Fristverlängerung der bisherigen Zertifikate unter genau definierten Bedingungen möglich. Dies ist u. a. von der Art der Produkte, von der bisherigen Gültigkeit der Zertifikate sowie auch von dem aktuellen Status der Umstellung auf die MDR möglich (im letzten Fall nachweislich durch einen Vertrag mit der Benannten Stelle gemäß Anhang VII Abschnitt 4.3 Unterabsatz 2 MDR).

In diesem Übergangsprozess prüfen wir mit besonderer Sorgfalt die Gültigkeit der Zertifikate der Hersteller unserer Produkte. Im Falle, dass ein Zertifikat nicht mehr gültig ist, ist eine weitere Belieferung an uns nur möglich, wenn die o.g. Bedingungen ohne Einschränkung erfüllt sind. Wir bestätigen somit, dass die Hersteller unserer Produkte ein aktuell gültiges Zertifikat besitzen oder die Bedingungen für die Fristverlängerung nach MDR erfüllen.

Haben Sie Fragen diesbezüglich oder benötigen detailliertere Informationen, wenden Sie sich gerne an uns.

Kay Kruse

Geschäftsführung/Sicherheitsbeauftragter

GHC German Health Care GmbH
Birklück 15
D- 24999 Wees

Telefon: +49 (0) 4631 / 572 00 0
Telefax: +49 (0) 4631 / 572 00 19
E-Mail: info@ghc-medical.de

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Benannt durch/Designated by
Zentralstelle der Länder für Gesundheitsschutz bei Arzneimitteln und Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08

Product Service

EC Certificate

Full Quality Assurance System
Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 036336 0054 Rev. 03

Manufacturer:

Zhejiang Kindly Medical Devices Co., Ltd.
No.758, 5th Binhai Road
Binhai Industrial Park, Longwan District
325025 Wenzhou, Zhejiang Province
PEOPLE'S REPUBLIC OF CHINA

Product Category(ies):

Disposable Needles, Scalp Vein Sets, Blood-Collecting Needles, Huber Needles, Fistula Needles, Anaesthesia Needles, Dental Needles for Single Use, Sterile I.V. catheter for single use, Disposable Insulin Pen Needle, Sterile Biopsy Needles for single use, Sterile Percutaneous Vertebroplasty Kit for single use, Sterile Irrigation Needles for Single Use, Safety Needles,Safety Scalp Vein Sets, Safety Blood-Collecting Needles, Safety I.V. Catheter for Single Use, Safety Fistula Needles, Luer Adapter, Safety Blood Lancet, Syringes, Infusion Sets, Transfusion Sets, Burette-Type Infusion Sets, Sterile Intravascular Catheter Introducer for Single Use, Sterile Syringes for Insulin for Single Use, Sterile Disinfecting Cap for Single Use.

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II.
This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G1 036336 0054 Rev. 03

Report No.: BJ20081201

Valid from: 2020-10-27
Valid until: 2024-05-26

Date, 2020-10-27

Christoph Dicks
Head of Certification/Notified Body

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany